



Lumis[™] series
Noninvasive ventilators



NIV Technologies: Lumis[™] handbook

Table of contents

Lumis Technologies at a glance3

Accessing the Lumis clinical menu.....4

Lumis ventilation modes5

Advanced setting to optimise comfort6

Rise Time6

Trigger and Cycle Sensitivities.....7

Ti Control.....8

Setting Ti Min and Max9

Lumis comfort features.....10

Ramp and Ramp Down10

Smart Start10

Vsync.....11

Climate Control.....12

Lumis 150 ST-A alarm modules.....13

NIV Titration and troubleshooting14

Intelligent Volume Assured Pressure Support (iVAPS)16

iBR.....17

AutoEPAP18

iVAPS titration protocol for Lumis19

Manual titration using Target Va Calculator19

Automated titration using Learn Targets20

Full titration using EasyCare Tx21

Lumis series accessories.....23

Lumis Oximetry set-up procedure.....25

Connecting the Lumis Oximetry Module25

Lumis RPS II set-up procedure26

Lumis reprocessing guidelines27

Reprocessing the air tubing27

Reprocessing the humidifier and air outlet.....29

Reassembling.....31

Lumis technologies at a glance

Therapy modes	Lumis 100 VPAP S (28012)	Lumis 150 VPAP ST (28124)	Lumis 150 VPAP ST-A (28223)
CPAP (Continuous Positive Airway Pressure)	•	•	•
S (Spontaneous)	•	•	•
T (Timed)		•	•
ST (Spontaneous-Timed) with optional iBR		•	•
PAC (Pressure Assist-Control)		•	•
iVAPS (with iBR) and optional AutoEPAP		•	•
Features			
Alarms			•
Vsync	•	•	•
TiControl	•	•	•
Trigger and cycle	•	•	•
EPR™			
Ramp	•	•	•
Ramp Down	•	•	•
Climate Control Auto	•	•	•
Essentials mode	•	•	•
Data management			
Built-in wireless connectivity*	•	•	•
AirView	•	•	•
AirView's remote assist	•	•	•
ResScan	•	•	•


*Wireless communication depends on network availability.

2 NIV Technologies handbook

Accessing the Lumis clinical menu



Hold down the home button and dial together for three second to unlock the Clinical Menu



Lumis ventilation modes

CPAP	T
CPAP (continuous positive airway pressure) mode delivers one fixed level of pressure during both the patient’s inspiratory and expiratory phase.	T (timed) mode delivers pressure at a fixed breath rate, and a fixed inspiration and expiration time set by the clinician, regardless of the patient’s effort.
S	P(A)C*
S (spontaneous) mode senses both the patient’s inspiratory and expiratory phase, and delivers a preset level of pressure during each to support the patient’s spontaneous breathing.	P(A)C (pressure assist control) is a mandatory pressure mode where the inspiration time is preset, with inspiration initiated spontaneously or time triggered should the patient’s respiratory rate fall below the backup rate.
ST	iVAPS
ST (spontaneous timed) mode delivers a preset level of pressure to support the patient’s spontaneous breathing while providing additional breaths should the patient’s respiratory rate fall below the backup rate.	iVAPS (intelligent volume-assured pressure support) mode automatically adjusts the pressure delivered to meet a patient’s target alveolar ventilation.

The common adjustable parameters for different modes in the Lumis device are shown below.


Parameter	S	ST	T	PAC	iVAPS	CPAP
Set Pressure						•
IPAP	•	•	•	•		
EPAP	•	•	•	•	•	
Min PS					•	
Max PS					•	
Mini EPAP*					•	
Max EPAP*					•	
Resp. Rate			•			
Backup Rate		•		•		
Target Pt Rate**		•			•	
Target Va					•	
Ti			•	•		
Ti Max	•	•			•	
Ti Min	•	•			•	
Rise Time	•	•	•	•	•	
Trigger	•	•		•	•	
Cycle	•	•			•	
Height					•	

* Only available when AutoEPAP is enabled ** Only available when iBR is enabled in ST mode

Advanced settings to optimise comfort


Rise Time

Rise Time is the time it takes for inspiratory pressure to rise to the prescribed level. A controllable Rise Time enhances your patient’s breathing effort, comfort and synchrony. Set in milliseconds with a settable range of MIN/150 -900ms. Reports of the air delivery being ‘too hard’, ‘too fast’, ‘too strong’ or ‘intolerable’ may indicate that the Rise Time may need to be lengthened. Reports of the air delivery being ‘too slow’ or ‘not enough air’ may indicate a need for a shorter Rise Time.



Trigger and Cycle Sensitivities

Trigger and Cycle customises the beginning and end of each breath to reflect your patient’s condition and is designed to maintain patient-ventilator synchrony, and reduce the effort required to breathe.



What does Trigger and Cycle do?

ResMed ventilators employ flow triggering and cycling.

- Under normal conditions, the device triggers (initiates IPAP) and cycles (changes to EPAP) as it detects changes in patient flow.
- Trigger settings can be used to adjust the amount of flow required to trigger inspiration.
- Cycle settings can be used to adjust the percentage of peak inspiratory flow required to cycle to expiration.

What makes Trigger and Cycle special?


- **Fine-tuning and customisation:** five adjustable breath detection sensitivity settings are available and can be used to optimise the detection of the beginning and end of the breath and customise the technology to meet your patient’s unique needs.
- **Flexibility:** a wide range of settings enables you to personalise therapy, comfort and synchrony across a wide range of pathologies, from paediatric to adult and from restrictive to obstructive.
- **Synchronisation:** functions in parallel with ResMed’s Vsync automatic leak management technology and is designed to enhance breath detection and optimise patient-device synchronisation.

Trigger and cycle sensitivity table

Trigger settings	Very Low	Low	Medium	High	Very High
	Less sensitive/Harder to Trigger		Default	More sensitive/Easier to Trigger	
Cycle settings	Very Low	Low	Medium	High	Very High
	Less sensitive/Harder to Cycle		Default	More sensitive/Easier to Cycle	



TiControl maintains the ideal inspiratory cycle for your patient’s needs. Minimum and maximum inspiratory time limits can be set above and below a patient’s spontaneous inspiration time, for enhanced patient–ventilator synchrony.



What does TiControl do?

TiControl tailors your patient’s breathing to suit their condition by:

- enabling you to set upper and lower limits on inspiratory time
- enabling you to manage spontaneous flow and set ideal levels that match your patient’s unique needs
- reacting swiftly and effectively to correct unstable breathing.

What makes TiControl special?

- **Reassuring responsiveness:** provides timely intervention if your patient’s breathing cycle becomes spontaneous and shifts away from its registered inspiration phase.
- **Personalised fine-tuning:** allows you to limit minimum and maximum inspiratory periods to suit your patient’s unique situation. For example, you can deliver a shortened Ti Max phase if your COPD patient has a restricted respiratory flow and difficulties exhaling.

This could help to prevent late cycling to expiratory pressure and provide a better match with their ideal inspiratory time.

- **Patient-device synchronisation:** works in combination with Vsync technology and is designed to offer extra assurance, control over your patient’s inspiratory time and maintain outstanding patient-device synchronisation.

TiControl titration table

Note: This table is meant to be used as a guide. The clinician shall exercise due caution and judgement, while ensuring the patient is comfortable.

Setting Ti Min and Max

1. Instruct the patient to breathe normally while either lying down or sitting in an upright position.
2. Observe and record the number of breaths the patient takes per minute (BPM).
3. Consider the patient’s respiratory condition and refer to the appropriate table for that condition i.e. Restrictive, Obstructive or Normal.
4. Match the patient’s spontaneous BPM to the BPM on the left hand side of the table.
5. Input the corresponding Ti Min and Max setting.

Restrictive States		
BPM	Ti Min: (ratio 1:3)	Ti Max: (ratio 1:1)
30	0.5	1.0
29	0.5	1.0
28	0.5	1.1
27	0.6	1.1
26	0.6	1.2
25	0.6	1.2
24	0.6	1.3
23	0.7	1.3
22	0.7	1.4
21	0.7	1.4
20	0.8	1.5
19	0.8	1.6
18	0.8	1.7
17	0.9	1.8
16	0.9	1.9
15	1.0	2.0
14	1.1	2.1
13	1.2	2.3
12	1.3	2.5
Examples of Restrictive states are; Neuro-Muscular Disease and Duchenne Muscular Dystrophy		

Obstructive States		
BPM	Ti Min: (leave on default setting - 0.3 sec)	Ti Max: (ratio 1:2)
30	0.3	0.7
29	0.3	0.7
28	0.3	0.7
27	0.3	0.7
26	0.3	0.8
25	0.3	0.8
24	0.3	0.8
23	0.3	0.9
22	0.3	0.9
21	0.3	0.9
20	0.3	1.0
19	0.3	1.0
18	0.3	1.1
17	0.3	1.2
16	0.3	1.2
15	0.3	1.3
14	0.3	1.4
13	0.3	1.5
12	0.3	1.7
Examples of Obstructive states are; Chronic Obstructive Pulmonary Disease (COPD)		

Normal States		
BPM	Ti Min: (leave on default - 0.3 sec or use a 1:2 ratio as below)	Ti Max: (ratio 1:1)
30	0.7	1.0
29	0.7	1.0
28	0.7	1.1
27	0.7	1.1
26	0.8	1.2
25	0.8	1.2
24	0.8	1.3
23	0.9	1.3
22	0.9	1.4
21	0.9	1.4
20	1.0	1.5
19	1.0	1.6
18	1.1	1.7
17	1.2	1.8
16	1.2	1.9
15	1.3	2.0
14	1.4	2.1
13	1.5	2.3
12	1.7	2.5
Examples of Normal states are; Obesity Hypoventilation Syndrome (OHS) & Post-Operative Applications		


Lumis comfort features

Ramp and Ramp Down

Ramp gradually and comfortably increases the pressure until the prescribed therapy level is reached.

Ramp Down eases your patient into spontaneous breathing by gradually reducing the pressure.

Ramp Down gradually decreases the therapeutic pressure until Start EPAP is reached. Lumis decreases the pressure over a fixed 15-minute period then continues in CPAP mode.




SmartStart

With the SmartStart feature enabled, the device will start automatically when the patient breathes into the mask and then stop automatically when the patient removes the mask.


With SmartStart enabled, the start/stop button will still function if pressed.

With the High Leak alarm enabled (in Lumis ST-A device only), the SmartStart function will be automatically disabled.



Vsync is designed to ensure reliable, patient-device synchrony and rapidly, accurately and automatically detect and compensate for leaks. This essential technology underpins excellence in NIV ventilation.

Vsync leak management algorithm



What does Vsync do?

Vsync supports successful ventilation by:


- rapidly detecting and responding to leaks
- accurately compensating for dynamic, fast-changing and unstable leaks
- effectively restoring patient-device synchrony, usually within 2-3 breaths.

What makes Vsync special?

- **Rapid response:** shifts to a 2-second detection and re-synchronisation phase as soon as a leak is detected. As soon as patient-device synchrony is restored, it reverts to the default 10-second mode.
- **Smart algorithm:** recognises the difference between patient inhalation and leak. This reduces the risk that the ventilator will cycle your patient's breath ineffectively or mis-trigger from EPAP to IPAP, out of sync with actual effort.
- **Reassuring insight:** reliably measures essential patient information including tidal volume, minute ventilation and respiratory rate.
- **Patient-device synchronisation:** complements TiControl technology and is designed to maintain outstanding patient-device synchronisation, even in cases of significant leak.



Advanced humidification that adapts to changing ambient conditions.



What does Climate Control do?

Climate Control is a comfort feature designed to automatically adjust to:

- changes in ambient room temperature and humidity,
- changes in flow caused by pressure changes,
- changes in flow due to mask or mouth leak.

What makes Climate Control special?

- **Comfort:** feedback from five sensors intelligently adjusts the humidity to deal with changes in ambient conditions.
- **Flexibility:** in Manual mode, the tube temperature and humidity level can be set independently, so your patient can control the comfort of their therapeutic experience.
- **Simplicity:** in Auto mode, temperature and humidity are pre-set at levels designed to deliver increased comfort, providing your patient with immediate access to comfortable humidification at the touch of a button.

Lumis 150 ST-A alarm modules

Alarm types

All the alarms on the device are classified as medium priority.

Fixed alarms

The alarms pre-set for the device are:

- Power fail
- Blocked tube*
- Tube disconnected
- System fault (system error).

*Only triggered reliably for pressures above 10 cm H₂O.

Adjustable alarms

Alarms that can be set are:

- High leak
- Non-vented mask
- Low minute ventilation
- Apnoea
- Low SpO₂ (when oximeter connected).

The alarms listed above can also be catergorised as:


- Clinical alarms—Low minute ventilation, Apnoea, Low SpO₂
- Patient circuit alarms—Blocked tube, Tube disconnected, High leak, Non-vented mask
- System alarms—Power fail, System fault.

Alarm	Description	Range
High Leak*	Enable / disable the High Leak alarm. Activates when a leak >40 L/min (0.7 L/sec) occurs for >10–30 seconds.	On / Off
Non-Vented Mask	Enable / disable the Non-Vented Mask alarm. Activates within 20–40 seconds when a non-vented mask is attached during therapy. Note: Use of supplemental oxygen with a vented	On / Off
Low MV*	Sets the Low Minute Ventilation alarm. Activates within 20–40 seconds after the measured level remains below the set limit. Note: Alarm may not trigger reliably when using a P10 mask.	Off / 1–10 L/min, 1 L/min increments
Low SpO ₂	Sets the Low SpO ₂ alarm. Activates when the SpO ₂ value is below the set value for 20–40 seconds. Only functional when an oximeter is connected. When the Low SpO ₂ alarm is set, the Oximeter sensor disconnected alarm and Oximeter sensor failure alarm are also enabled.	Off / 70–95%, 1 % increments
Apnoea	Sets the Apnoea alarm. Activates when there is no inspiratory trigger (either patient or machine) detected from the previous inspiration for the set interval. Two consecutive inspirations (either patient or machine triggered) reset the Apnoea alarm.	Off / 10–60 sec, 1 sec increments
Alarm Volume	Sets the alarm volume.	Low / Medium / High

*When enabled, SmartStart is automatically disabled.


NIV Titration and troubleshooting

NIV Initiation Protocol for Chronic Respiratory Insufficiency or Failure



Intelligent Volume Assured Pressure Support (iVAPS)

iVAPS is ResMed's unique volume-assurance mode designed for adults with chronic respiratory failure whose condition is likely to change during therapy. It is designed to maintain a preset target alveolar ventilation by monitoring delivered ventilation, adjusting the pressure support and providing an intelligent backup breath automatically.



What does iVAPS do?

iVAPS automatically:

- monitors the presence and volume of air flow
- delivers the volume of air your patient needs for their next breath.


What makes iVAPS special?

- **Targets alveolar ventilation:** iVAPS enables improved control of ventilation by taking account of tidal volumes, respiratory rates and the anatomical dead space in your patient's airways.
- **Responds to change:** reacts rapidly but gently to maintain your patient's unique alveolar ventilation target and stabilise blood gases, even when their respiratory rate changes during the night.
- **Learns and adapts:** the Learn Targets feature uses information about your patient's alveolar ventilation and respiratory rate to automatically set targets and streamline titration. Targets can also be entered manually.
- **Improves adherence:** compared to standard pressure support ventilation, iVAPS helps patients adhere to therapy for 60 minutes longer per session.



iBR

iBR (intelligent Backup Rate) is designed to enhance the conventional approach to backup breaths. It maximises your patient's opportunities for spontaneous breathing by delivering support only when it's needed and tailoring that support to meet their real needs.



What does iBR do?

iBR supports spontaneous breathing by:

- providing backup breaths within two boundaries: target patient rate and background frequency which is two-thirds of the target rate
- responding only when the patient needs support
- gradually adjusting the backup breath from the lower limit (background frequency) to the patient's own rate when an apnoea or lack of effort is detected.


What makes iBR special?

- **Interprets information:** iBR will provide backup breaths if there's an apnoea or lack of effort, but won't provide unnecessary support for pauses caused by a cough or a sigh.
- **Reacts swiftly:** designed to safely and quickly return your patient to their target rate, typically within 4-5 breaths. A single spontaneous breath resets the iBR and re-establishes the background frequency until backup support is next required.
- **Time-saving personalisation:** automatically determines the most suitable iBR for your patient based on their spontaneous stable awake rate, giving you peace of mind and ensuring your patient receives personalised ventilation.


AutoEPAP

AutoEPAP automatically adjusts expiratory positive airway pressure (EPAP) to maintain an open upper airway. It aims to prevent full and partial obstructions from occurring and automatically adjusts expiratory pressure settings to address any obstructions that do occur.

The following inspiratory flow shapes are indicative of partial upper airway obstruction.



Severe to complete reduction of airflow is indicative of a complete upper airway obstruction.




What does AutoEPAP do?

AutoEPAP maintains an open upper airway by:

- identifying full and partial obstructions
- calculating the severity of the event
- providing the minimum level of EPAP required to meet your patient's needs,
- gradually decreasing the pressure once the obstructive event has been addressed.

What makes AutoEPAP special?


- **Manages complexity:** finely balances the multiple parameters required to ventilate adults with chronic respiratory failure. AutoEPAP maintains upper airway patency even if target expiratory pressure fluctuates due to variations in your patient's primary condition, body position, weight or medication.
- **Swift, automatic response:** automatically adjusts expiratory pressure within Min EPAP and Max EPAP settings to address full and partial obstructions, then re-establishes the right balance between upper airway patency and breathing comfort.



iVAPS titration protocol for Lumis

Manual titration using target Va Calculator

- 1 Set up the circuit configuration and default settings
- 2 Select iVAPS mode
- 3 Enter the patient's height



- 4 Set the target patient rate (Target Pt Rate)
- 5 Determine the target alveolar ventilation (Target Va) with the help of the relevant tidal volume (Vt) or minute ventilation (MV)
- 6 Check the calculated values
- 7 Start ventilation
- 8 Re-evaluate the settings

- Target Pt Rate should be equivalent to the patient's spontaneous respiratory rate (RR) during ventilation.
- The patient's typical values (RR, MV and Vt) can be viewed on the monitoring screens, or in ResScan™ – ResMed's data management software. These values should be adopted from the stable breathing phase: spontaneous triggering with no desaturation.

Automated titration using Learn Targets

1

Set up the circuit configuration and default settings

2

Select iVAPS mode

3

Enter the patient's height

4

Select Learn Targets

Learn Targets

< Home

Learning... 20:00

Height 175cm

EPAP 8.0

PS 6.0

< Stop Learning

Min PS 3.0

Lumis

5

Enter preliminary PEEP/EPAP (e.g., 4–6 cm H₂O). If AutoEPAP is on, the EPAP setting for Learn Targets is Min EPAP

6

Adjust pressure support (PS) (e.g., 3–4 cm H₂O)

7

Wait until the patient's breathing appears stable (a minimum of 20 minutes)

8

Accept stable Target Va and RR. Ventilation will start automatically

9

Re-evaluate the settings

- During Learn Targets, the patient should be awake, seated and show stable breathing.
- There is no backup rate during Learn Targets.
- Pressure support should be set to a clinically relevant level depending on the patient's condition. For patients who maintain adequate daytime ventilation, i.e., patients with obesity hyperventilation syndrome (OHS), the goal is to capture the resting daytime ventilation with minimal pressure support. Patients with significant daytime hypercapnia may require greater levels of pressure support (i.e., 6–10 cm H₂O).

Full titration using EasyCare Tx

A

iVAPS titration by entering the target values directly

1

Set up the circuit configuration and default settings

2

Select iVAPS mode

3

Set

Height (e.g., 180 cm)
Target Pt Rate (e.g., 15 bpm)
Target Va (e.g., 6 mL/kg IBW)
EPAP (e.g., 5 cm H₂O)
Min PS (e.g., 4 cm H₂O)
Max PS (e.g., 20 cm H₂O)

iVAPS Settings

Height : 180

Target Patient Rate : 10.0

Target Alve Ventilation : 5.3

EPAP : 5.0

Min PS : 4.0

Max PS : 20.0

Ti Max : 1.5

Ti Min : 0.5

Rise Time : 300

Fall Time :

Trigger : Medium

Cycle : Medium

Calculated values

MV

Vt

Vt/kg

6.6

0.66

9.0

OK

Cancel

4

Start ventilation

5

Check flow, SpO₂ and CO₂. Adjust settings if necessary

B

Titration with transition from bilevel mode to iVAPS mode

1

Titrate in bilevel mode and check if MV, Vt and RR are stable

2

Select iVAPS mode

3

Set

Height (e.g., 180 cm)
Min PS (e.g., 4 cm H₂O)
Max PS (e.g., 20 cm H₂O)

Switch Bilevel (std) to iVAPS

From Bilevel (std)

Mode : S

IPAP : 4

EPAP : 4

Ti Min : 0.5

Ti Max : 1.5

Rise Time : 300

Trigger : Medium

Cycle : Medium

To iVAPS

Height : 180

Target Patient Rate : 28

Target Alve Ventilation : 1.0

EPAP : 4.0

Min PS : 0.0

Max PS : 20.0

Ti Min : 0.5

Ti Max : 1.4

Rise Time : 250

Trigger : Medium

Cycle : Medium

Measured Patient Values

Minute Ventilation : 4.1

Respiratory Rate : 29

Calculated values

MV

Vt

Vt/kg

4.1

0.15

2.0

OK

Cancel

4

Start ventilation

5

Check flow, SpO₂ and CO₂. Adjust settings if necessary

- In response to obstructive apnoeas, hypopnoeas, snoring and flow limitation, increase EPAP incrementally, e.g., by ≥1 cm H₂O every five minutes. Alternatively, select AutoEPAP: a setting designed to automatically maintain upper airway patency. The AutoEPAP algorithm does not address any other titration target, such as lung recruitment, to improve oxygenation or offset intrinsic PEEP. Min EPAP should be set to treat lower airway conditions.
- To address desaturation in the absence of upper airway obstruction, increase Target Va incrementally, e.g., by 0.3 L/min every five minutes.
- The Target Pt Rate should be equivalent to the patient's spontaneous respiratory rate.

Frequently asked questions

Who can benefit from iVAPS and AutoEPAP?

Patients (weighing 30 kg or more) with chronic respiratory failure or insufficiency, e.g., neuromuscular disease (NMD), obesity hypoventilation syndrome (OHS) and chronic obstructive pulmonary disease (COPD).

What are the recommended default values?

	Min PS	Max PS	EPAP	Min EPAP	Max EPAP
Obstructive	4 cm H ₂ O	20 cm H ₂ O	5 cm H ₂ O	5 cm H ₂ O	15 cm H ₂ O
Restrictive	4 cm H ₂ O	20 cm H ₂ O	5 cm H ₂ O	5 cm H ₂ O	15 cm H ₂ O
OHS	4 cm H ₂ O	18 cm H ₂ O	8 cm H ₂ O	8 cm H ₂ O	15 cm H ₂ O
Normal	2 cm H ₂ O	20 cm H ₂ O	5 cm H ₂ O	5 cm H ₂ O	15 cm H ₂ O

Some clinicians may choose to increase Min PS for patients who have been established in bilevel modes.

Example:

Treated in ST mode	Transition to iVAPS mode
IPAP: 20 cm H ₂ O	Min PS: 12 cm H ₂ O
EPAP: 5 cm H ₂ O	
PS: 15 cm H ₂ O	Max PS: 20 cm H ₂ O

What is the typical Target Va value?

The default value for the Target Va is 5.2 L/min.

	Target Va
COPD	<11 L/min
Non-COPD	<7 L/min

Does the Target Va need to be adjusted during follow-up?

The Target Va needs to be assessed and adjusted just like IPAP and EPAP in bilevel mode.

How can PaCO₂ be reduced further?

To decrease PaCO₂ by 10%, increase Target Va by 10%.

PaCO₂: **55** mm Hg → **50** mm Hg
Target Va: **6** L/min → **6.6** L/min

What should be considered when setting PEEP/AutoEPAP?

PEEP (or EPAP) should be set to a level according to each of your patients' needs. Examples of its use include:

- Recruiting/maintaining lung volume, aiding gas exchange, improving lung compliance or minimising lung injury.
- In obstructive lung disease – to offset intrinsic PEEP, to reduce work of breathing and/or to maximise effective triggering.
- To splint open or stabilise the upper airway during sleep. For nocturnal ventilation, a starting pressure of 5 cm H₂O is recommended. An additional 1–3 cm H₂O may be helpful if there is any indication of symptoms such as daytime sleepiness, obesity, an anatomically narrow airway (including NMD), or pressure support frequently reaching the maximum level.

It is important to note that AutoEPAP is designed only to maintain upper airway patency by automatically adjusting pressure in response to flow limitation or obstruction of the upper airway. To treat lower airway conditions, the Min EPAP can be set to help improve lung recruitment or offset intrinsic PEEP.

What should be assessed at follow-up?

Consider the following:

- Does the patient sleep well?
- Is leak present?
- What is the level of SpO₂?

Review the change of pressure via ResScan

If the pressure support remains high:

- Is Target Va too high?
- Is Max PS too high?
- Is EPAP or Min EPAP (when AutoEPAP is on) too low?

Check in ResScan for upper airway obstruction

- Are there episodes of desaturation?
- Does the flow graph show small flattened breaths?
- Is there high pressure support without associated respiratory flow?
- If yes, adjust EPAP or Min EPAP (when AutoEPAP is on).

Lumis™ series accessories

Product image	Description	Order code	MRRP
	ClimateLine™ Heated Tube	37296	\$99.00
	Standard Tubing	14994	\$18.00
	HumidAir™ Humidifier Tub (re-usable)	37300	\$129.95
	Air Outlet	37310	\$9.95
	90W PSU	37347	\$49.95
	USB Module	37301	\$69.95
	SD Card (1 pk) SD Card (12 pk)	37334 37332	\$19.95 \$179.95
	Standard Filter (2 pk) Standard Filter (12 pk)	36852 36852	\$5.90 \$29.90
	Hypoallergenic (2 pk) Hypoallergenic (12 pk)	36856 36857	\$7.50 \$34.90
	Oximeter Module	37302	\$120.00
	XPOD LP Oximeter	22374	\$750.00
	XPOD LP Oximeter Clip	22371	\$6.00

Lumis™ series accessories

Product image	Description	Order code	MRRP
	NONIN Oximeter Reusable Soft Sensor (8000SS) - Small (1m)	70567	\$300.00
	NONIN Oximeter Reusable Soft Sensor (8000SM) - Medium (1m)	70568	\$300.00
	NONIN Oximeter Reusable Soft Sensor (8000SL) - Large (1m)	70413	\$300.00
	NONIN Oximeter Reusable Finger Clip Sensor - 8000AA (1m)	631803	\$300.00
	Homecare Stand without Interface Top Plate	27963	\$700.00
IMAGE NOT AVAILABLE	Air10/Lumis Series Top Plate	27964	\$100.00
	Homecare Stand IV Pole Kit	27948	\$80.00
	RPS II External Battery Pack (includes Battery, Power Supply Unit adapter and Carry Bag)(Note: S9 Power Supply Unit 90W & DC Cable is not included and must be ordered separately.	24921	\$549.00
	ResMed Power Station II (Battery only)	24926	\$520.00
	RPS II - DC Cable Lumis Note: RPS II Lumis PSU Adapter is also required	37343	\$43.00
	RPS II - Lumis PSU Adapter Note: RPS II Lumis DC Cable is also required	37342	\$41.00

Lumis oximetry set-up procedure



Lumis/
Air10
Oximetry
Module

NONIN
XPOD
LP
Oximeter

NONIN
Soft Sensor


Items required:

- 37302 – Lumis/Air10 Oximetry Module/ Adapter
- 22374 – NONIN XPOD 3012 LP Oximeter
- 22371 – XPOD LP Oximeter Clip (optional)
- 70568 – NONIN 8000SM Soft Sensor – Reusable (Medium)

Or alternatively, the below kit is available.

- 22370 - Oximetry Set with Reusable Sensor - includes 1 x XPOD LP Oximeter, 1 x XPOD LP Oximeter Clip, 1 x NONIN Oximeter Reusable Soft Sensor (8000SM)

Connecting the Lumis Oximetry Module




Step 1

Prepare the Lumis with the Oximeter module attached and with the XPOD and finger sensor connected. Attach the finger sensor to the patient. Connect power to the Lumis. There should be a brief message displayed on the Lumis LCD as shown above – Oximetry SX489-0200.



Step 2

Make sure the Lumis machine has the SD card installed or you will see this message. If the card has been removed then power down the Lumis, insert the SD card and then reapply the power.



Step 3

If there is a problem with the finger sensor then you will see this message.

Lumis™ RPS II set-up procedure



RPS II Battery Pack

RPS II Lumis/ Air 10 PSU Adapter

RPS II Lumis/ Air 10 DC Cable

Items required:

- 24921 – RPS II Battery Pack (includes battery, PSU adapter, and carry bag)
- OR
- 24926 – RPS II Battery only
- AND
- 37343 – RPS II Lumis/Air10 DC Cable
- 37342 – RPS II Lumis/Air10 PSU Adapter

Lumis reprocessing guidelines

This guide is intended for multipatient use of the Lumis™ devices.

When the device is used for multiple patients, for example, in a sleep lab, clinic, hospital or at a health care provider, the cleanable humidifier, air outlet and air tubing should be reprocessed between each patient use.

If the cleanable humidifier or the air tubing are being used for a single user in the home, refer to the cleaning instructions in the User Guide.


Described here are ResMed’s recommended and validated procedures for cleaning and disinfecting the cleanable humidifier, air outlet and air tubing. However, the steps for disinfection vary regionally and each healthcare facility should consult its own procedures before carrying out those within this guide.

- WARNING**
- **ResMed cannot give any assurance that deviations from the procedures listed in this guide, and their effect on the performance of the product, will be acceptable.**
 - **When using detergents, disinfectants or sterilisation agents, always follow the manufacturer's instructions.**
 - **Beware of electrocution. Do not immerse the device, power supply or power cord in water. If liquids are spilled into or onto the device, unplug the device and let the parts dry. Always unplug the device before cleaning and make sure that all parts are dry before plugging it back in.**


- ### Surface disinfection
1. Wipe the exterior of the device including display, externally accessible ports, side cover, power supply unit and accessories with a disposable cloth and mild detergent or alcohol disinfectant (see list below).
 2. Remove any excess disinfectant with a disposable dry cloth.
- Agents recommended for surface disinfection and cleaning:
- Warm water and mild detergent eg, Teepol™ multipurpose detergent
 - Window cleaner or other premixed surface detergent
 - Methyl alcohol solution
 - 70% Ethyl alcohol solution
 - 70-90% Isopropanol solution
 - 10% Bleach solution
 - Isopropyl wipes
 - CaviCide™
 - Mikrozid®
 - Actichlor™ Plus
 - Terralin®.
- Note: Agents may not be available in all regions.

Reprocessing the air tubing

Disconnecting



Hold the cuff of the air tubing and gently pull it away from the device



Hold both the cuff of the air tubing and the swivel of the mask, then gently pull apart.



Decontaminating

Before the disinfection process, each component must be cleaned and rinsed so no visible contamination is present.

- 1. Clean all components with a soft bristled brush for one minute while soaking in detergent solution (see table below). Pay particular attention to all crevices and cavities.
- 2. Run the detergent solution through the air tubing repeatedly until no contamination is visible.
- 3. Thoroughly rinse each component according to the detergent manufacturer's instructions

ResMed has tested the following detergents according to the manufacturer’s instructions:

Detergent	Water temperature	SlimLine/Standard	ClimateLine Air	ClimateLine Air Oxy
Alconox™ (diluted at 1%)	Hot water (approx 60°C) Warm water (approx 45 to 60°C) Room temperature water (approx 21°C)	•	•	•
Neodisher MediZym™ (diluted at 2.0%)	Warm water (approx 45 to 60°C)	•		
Gigazyme® (diluted at 1.0%)	Room temperature water (approx 21°C)		•	

Disinfecting

In the procedures below, only one disinfection process needs to be performed.

High level thermal

Part	Validated number of cycles
	Hot water: 75°C for 30 minutes OR 70°C for 100 minutes.
SlimLine	20
ClimateLineAir	26
ClimateLineAir Oxy	20
Standard	100

- 1. Soak the air tubing in a commercially available solution of a chemical sterilant. Take care that no air bubbles are trapped inside the air tubing.
- 2. Thoroughly rinse the air tubing in drinking quality water (five litres per assembly) by immersing it completely for a minimum of one minute in duration.
- 3. Repeat the rinse procedure two additional times using fresh water for a total of three rinses.

High level chemical

Part	Validated number of cycles		
	CIDEX® OPA Ortho-phthalaldehyde 0.55% for 12 mins	Sterrad NX Sterrad 100S	Gigasept FF® 5% for 15 mins
SlimLine	100	-	-
ClimateLineAir	26	26	26
ClimateLineAir Oxy	20	20	-
Standard	100	-	-

- 1. Soak the air tubing in a commercially available solution of a chemical sterilant. Take care that no air bubbles are trapped inside the air tubing.
- 2. Thoroughly rinse the air tubing in drinking quality water (five litres per assembly) by immersing it completely for a minimum of one minute in duration.
- 3. Repeat the rinse procedure two additional times using fresh water for a total of three rinses.
- 4. Air dry out of direct sunlight and/or heat.

Inspecting

Perform a visual inspection of the air tubing. If any visible deterioration is apparent (holes, tears or cracks etc), the air tubing should be discarded and replaced. Slight discoloration may occur and is acceptable.

Reconnecting the air tubing

When the air tubing is dry, you can reconnect it to the device.

- 1. Connect the air tubing firmly to the air outlet located on the rear of the device.
- 2. Connect the free end of the air tubing firmly onto the assembled mask.

Packaging and storage


Store in a dry, dust-free environment away from direct sunlight.

Storage temperature: -20°C to 60°C.


Reprocessing the humidifier and air outlet

Disassembling


The following instructions provide guidance on how to correctly disassemble the cleanable humidifier and the air outlet.




Remove the humidifier from the device, open it and discard any remaining water.




Locate the air outlet on the inside of the device.




Hold the humidifier base and then fully open the humidifier lid and pull it away so that it easily detaches from the base.



Release the air outlet by pressing the clip located inside the device.



Remove the humidifier seal from the humidifier lid by pulling it away.



Remove the air outlet by pulling it out through the air outlet socket at the rear of the device.

Decontaminating

Before the disinfection process, each component must be cleaned and rinsed so no visible contamination is present.

- 1. Clean all components with a soft bristled brush for one minute while soaking in detergent solution (see table below). Pay particular attention to all crevices and cavities.
- 2. Thoroughly rinse each component according to the detergent manufacturer's instructions.

ResMed has tested the following detergents according to the manufacturer’s instructions:

Detergent	Water temperature	Cleanable humidifier	Air outlet
Alconox (diluted at 1%)	Hot water (approx 60°C) Warm water (approx 45 to 60°C) Room temperature water (approx 21°C)	•	•
Gigazyme (diluted at 1.0%)	Room temperature water (approx 21°C)	•	•

Disinfecting

In the procedures below, only one disinfection process needs to be performed.

High level thermal

Part	Validated number of cycles
	Hot water: 90°C for 1 minute OR 75°C for 30 minutes OR 70°C for 100 minutes. Due to specific regional requirements, ResMed cleanable humidifiers have been tested for disinfection (100 cycles) at 93°C for 10 minutes
Cleanable humidifier	130
Air outlet	130

- 1. Soak the disassembled components in a hot water bath at pasteurising temperature. Take care that no air bubbles are trapped against the components.
- 2. Air dry out of direct sunlight and/or heat.

High level chemical

Part	Validated number of cycles
	CIDEX OPA Ortho-phthalaldehyde 0.55% for 12 minutes Gigasept FF 5% for 15 minutes Sterrad NX Sterrad 100S
Cleanable humidifier	130
Air outlet	130

- 1. Soak the disassembled components in a commercially available solution of a chemical sterilant. Take care that no air bubbles are trapped against the components.
- 2. Thoroughly rinse the cleanable humidifier in drinking quality water (five litres per assembly) by immersing it completely for a minimum of one minute in duration.
- 3. Repeat the rinse procedure two additional times using fresh water for a total of three rinses.
- 4. Air dry out of direct sunlight and/or heat.


Inspecting

Perform a visual inspection of all components. If any visible deterioration is apparent (cracking, crazing, tears, etc), the humidifier should be discarded and replaced. Slight discoloration of the silicone components may occur and is acceptable.


Reassembling

The following instructions provide guidance on how to correctly reassemble the air outlet and the humidifier.



To reassemble the air outlet



Hold the air outlet with the seal pointing to the left and the clip pointing forward.



Make sure that the air outlet is correctly aligned and insert the air outlet into the socket. It will click in place.




Check if the air outlet is inserted correctly as shown.


Packaging and storage

Store in a dry, dust-free environment away from direct sunlight. Storage temperature: -20°C to 60°C.

To insert the humidifier seal:




Place the seal into the lid.




Press down along all edges of the seal until it is firmly in place.

To reassemble the humidifier lid:



Insert one side of the lid into the pivot hole of the base.



Slide the other side down the ridge until it clicks into place.

**ResMed Service Centre**

For all device repair requestes

Phone: 02 8884 2700

General email: service@resmed.com.au

ResMed Technical Services

For all technical troubleshooting assistance

Phone: 02 8884 2539

General email: techsupport@resmed.com.au

ResMed Australian Distribution Centre (ADC)

For all ordering & delivery needs

Phone: 1800 991 900 (option 1)

General email: resmedadc@resmed.com.au

**ResMed Ltd**

1 Elizabeth Macarthur Drive
Bella Vista NSW 2153 Australia

Lumis and TiControl are trademarks and/or registered trademarks of the ResMed family of companies. For patent information, see ResMed.com/ip.
Specifications may change without notice. © 2018 ResMed Ltd. 10110597/1 2018-02

